

IEC SYSTEM FOR CONFORMITY TESTING AND  
CERTIFICATION OF ELECTRICAL EQUIPMENT (IECEE)  
CB SCHEME

SYSTEME CEI D'ESSAIS DE CONFORMITE ET DE CERTIFICATION  
DES EQUIPEMENTS ELECTRIQUES (IECEE)  
METHODE OC

## CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Product  
*Produit*

Pulmonary Function Tester

Name and address of the applicant  
*Nom et adresse du demandeur*

Medical Graphics Corp  
350 Oak Grove Pky  
St Paul, MN 55127, USA

Name and address of the manufacturer  
*Nom et adresse du fabricant*

Medical Graphics Corp  
350 Oak Grove Pky  
St Paul, MN 55127, USA

Name and address of the factory  
*Nom et adresse de l'usine*

Medical Graphics Corp  
350 Oak Grove Pky  
St Paul, MN 55127, USA

Rating and principal characteristics  
*Valeurs nominales et caractéristiques principales*

24 V dc, 3 A (for reference only)

Trademark (if any)  
*Marque de fabrique (si elle existe)*

Not applicable

Model / Type Ref.  
*Ref. de type*

Express CCM, Express CPX

Additional information (if necessary)  
*Information complémentaire (si nécessaire)*

This CB Test Report comprises 7 enclosures. The CB Test Certificate was amended on June 18, 2009 to add component.

A sample of the product was tested and found to be in conformity with  
*Un échantillon de ce produit a été essayé et a été considéré conforme à la*

**PUBLICATION**

**EDITION**

**IEC 60601-1 (1988) Second Edition, with Amendment No. 1 (1991) and No. 2 (1995)** with the exception of: Clause 36, Electromagnetic Compatibility, Clause 48, Biocompatibility and Clause 52.1, Programmable Electronic Systems. Inclusive of CENELEC Common Modifications. See Test Report for National Differences.

E160574-A6-CB-1

as shown in the Test Report Ref. No. which forms part of this Certificate  
*comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat*

**This CB Test Certificate is issued by the National Certification Body**  
**Ce Certificat d'essai OC est établi par l'Organisme National de Certification**




**Underwriters  
Laboratories**

Underwriters Laboratories Inc. / Certification Programs Office, USA  
333 Pfingsten Road, Northbrook, IL 60062-2096  
United States of America  
TEL INT\* +1 847 664 3008, FAX INT\* +1 847 313 3008  
email: jolanta.m.wroblewska@us.ul.com

Date: Issued: 2007 May 23  
Amended: 2009 June 18 (Am. 1)

Signature:

  
Jolanta M. Wroblewska